K071260

## 510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

JUL 1 7 2007

JUL 17
3M ESPE AG
ESPE Platz
D-82229 Seefeld
Bavaria
Germany
9611385
Dr. Andreas Petermann,
Manager Regulatory Affairs
011-49-8152-700 1395
011-49-8152-700 1869
Andreas.Petermann@mmm.com
May 02, 2007
Hermes Bond 3
Resin tooth bonding agent
Dental Adhesive
K043043
K060684
K020256
K913965

Sinfony by 3M ESPE.....K992645

## Description for the Premarket Notification

Hermes Bond 3 is classified as Resin Tooth Bonding Agent (21 C.F.R. § 872.3200) because it is a device intended to be painted on the interior of a prepared cavity of a tooth to improve retention of restorative materials.

Hermes Bond 3 offers the advantages of a simplified bonding procedure, eliminating the need for a separate etching step. Thus it reduces both possible errors during application and post-operative sensitivity. Additionally, it saves the dentist valuable chair time. Hermes Bond 3 is intended to provide bonding between dentin/enamel and silorane (or oxirane) based composites. This is ensured by the formulation of Hermes Bond 3 which has especially been optimized for this purpose.

Hermes Bond 3 will be available in a two-vial version, one containing the Hermes Bond 3 Primer and one containing the Hermes Bond 3 Link (Bond).

To provide evidence for safety biocompatibility testing was carried out. The results show that Hermes Bond 3 is a safe device.

The comparison for chemistry, performance data and indications for use shows that Hermes Bond 3 is substantially equivalent to the predicate devices.

In summary, it can be concluded that safety and effectiveness requirements for Hermes Bond 3 are completely met.





Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

Dr. Desi W. Soegiarto Regulatory Affairs Specialist 3M ESPE AG Dental Products ESPE Platz Seefeld, Bavaria, GERMANY D-82229

JUL 17 2007

Re: K071260

Trade/Device Name: Hermes Bond 3

Regulation Number: 872.3200

Regulation Name: Resin Tooth Bonding Agent

Regulatory Class: II Product Code: KLE Dated: May 2, 2007 Received: May 7, 2007

## Dear Dr. Soegiarto:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <a href="http://www.fda.gov/cdrh/industry/support/index.html">http://www.fda.gov/cdrh/industry/support/index.html</a>.

Sincerely yours,

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital, Infection Control and Dental Devices Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known):

Device Name.	Hermes Bond 3	
Indications For Use:	Bonding between dentin/enamel and 3M ESPE silorane (or oxirane) based composites	
	•	
Prescription Use/ (Part 21 OFR 801 Subpart D)	AND/OR Over-The-Counter Use (21 CFR 801 Subpart C)	
(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)		
Concurrence of CDRF Office of Device Evaluation (ODE)		
(Division Sign-Off)		
Division of Anesthesiology, General Hospital Infection Control, Dental Devices		
-10(k) Number: <u>⟨</u> <u></u> <u></u> <u></u> <u> </u>	71260 Page 1 of	